

SPECTRA

Anaesthesia workstation with ventilator



Specifications:

- **FreshGas Flow:** 0-20 L/Min
- **Backbar:** Two Selectatec Vaporizers with Interlock Safety Mechanism
- **Absorber Capacity:** 1.5 Kg
- **Ventilation Modes:** VCV, PCV, ACV-V, ACV-P, Manual, Optional - SIMV-V, SIMV
- **LCD / TFT Display:** 7 Inch (Touch Screen)
- **Patient Type:** Adult / Pediatric
- **PEEP:** Electronically controlled
- **Trigger type:** Pressure, Flow
- **Tidal volume:** 10 - 1500 ml
- **Monitoring Parameter:** Ppeak, Vte, RR, MV, PEEP, FiO2
- **Waveform:** P-T, F-T, V-T
- **Ventilator Alarms:** Paw Max, Vte Max, Breathing Circuit Disconnect, O2 Pressure Low, Sensor Fault O2, RR High / Low, PEEP – High / Low, Vte – High / Low, MV – High / Low, Apnea, Ppeak – High / Low.
- **Event Logs Pre:** Up to 100 events.
- **Operative Checks:** Flow/Pressure Sensor Zero Calibration, 21% O2 Calibration, 100% O2 Calibration, Leak Test.
- **Power Supply:** 230 VAC / 50 Hz
- **Battery Backup:** 5 Hours

Available at:

XAVIER Med Pvt. Ltd.

16-2-741/9, Andhra Bank Colony,
Rishi Arcade, Moosarambagh, Hyderabad.
Ph: 8008895079, 8008895680