

COSMOS

Anaesthesia workstation with ventilator



Specifications:

FreshG as Flow: 0-20 L/Min

Backbar: Two Selectatec Vaporizers with Interlock

Absorber Capacity: 1.5 Kg

Ventilation Modes: VCV, PCV, ACV-V, ACV-P, SIMV-V, SIMV-P, Manual, PSV

LCD/TFT Display: 15 Inch (Touch Screen)

Patient Type: Adult/Pediatric

PEEP: Electronically controlled

Trigger Type: Pressure, Flow

Tidal Volume: 10 - 1500 ml

Monitoring Parameters: Ppeak, Vte, RR, MV, PEEP, FiO2

Waveform: P-T, F-T, V-T

Loop: P-V, F-V

Ventilator Alarms:

- Paw Max, Vte Max, Breathing Circuit Disconnect,
- O2 Pressure Low, Sensor Fault O2, RR High/Low,
- PEEP - High/Low, Vte - High/Low, mv - High/Low,
- Apnea, Ppeak - High/Low

Event Logs: Up to 100 events

Pre-Operative Checks: Flow/Pressure Sensor Zero Calibration, 21% O2 Calibration, 100% O2 Calibration, Leak Test

Power Supply: 230 VAC / 50 Hz

Battery Backup: 1.5 Hours

Patient Monitoring: ECG, SPO2, NIBP, Temp, EtCO2

Vital Alarms: ECG - Low/High, SPO2 - Low/High, NIBP - Low/High, Temp - Low/High

Vital Waveforms: ECG, SPO2, EtCO2 (Optional)

Vital Pre-Op Checks: EtCO2 (Zero Calibration), EtCO2 (Gas Compensation)

Available at:

XAVIER MED PVT. LTD.

16-2-741/9, Andhra Bank Colony,
Rishi Arcade, Moosarambagh, Hyderabad.
Ph: 8008895079, 8008895680